

Anklizer® II Walker

Please read the following instructions carefully and completely before use. Correct application is required to ensure proper function of the device.

Intended for single patient use.



WARNING: If your pain increases or persists or you experience any unusual reactions while using this product, consult your healthcare provider immediately.

APPLICATION INSTRUCTIONS

1. Unfasten straps, open liner, remove pads located inside boot and set aside.
2. While seated, place leg in brace and position heel at the back of the walker. Brace uprights should fall mid-line with the ankle.
3. Wrap liner around the leg and foot ensuring the ankle pad is positioned over the top of the foot. Secure hook to the liner.

NOTE: Liner and strap adjustment may be necessary as swelling patterns change throughout the rehabilitation process.

4. Position the uprights parallel to the leg per healthcare provider's recommendation.
5. Remove plastic sleeves from the uprights to expose hook.
6. Press hook on inside of the upright to the liner.
7. Beginning with the toe strap, thread straps through links and work up the leg by securing each strap.
8. Alternate directions of the top leg straps and secure to ensure the liner and uprights don't migrate.

NOTE: Pads may be inserted at either side of the ankle or at heel for additional comfort.

NOTE: Tall ankle walker has three leg straps and the low top ankle walker has one leg strap.

NOTE: Please refer to sew-in label on device for care instructions.

Please visit our website at www.birdcronin.com for additional product information or contact a customer care representative at **800-328-1095**.

NOTE: Every effort has been made to obtain the maximum compatibility of function, strength, durability and comfort with this product, however, there is no guarantee that injury will be prevented with its use.



Anklizer® II Walker Tall



Anklizer® II Walker Low Top

**NOT MADE WITH
natural rubber latex**