

**Cardinal Health™ MULTI-FLEX™  
Non-Sterile Latex Gloves**

# Non-sterile latex with hand-specific design



## Features and benefits

- Powder-free, natural rubber latex for comfort
- Independent thumb design helps reduce hand fatigue<sup>1</sup>
- Interlocking cuff design helps reduce roll-down
- Non-beaded cuff

Catalog No.	Size	Packaging
2Y1600	5½	100 gloves per polyethylene (PE) bag
2Y1601	6	
2Y1602	6½	
2Y1603	7	4 PE bags (2 left-hand + 2 right-hand) per case
2Y1604	7½	
2Y1605	8	400 gloves (200 pairs) per case
2Y1606	8½	
2Y1607	9	

## SPECIFICATIONS:

### Material

Latex, Powder-free

### Typical length

12 in. / 305 mm

### Color

Cream

### Cuff style

Interlocking

### Design

Hand specific with smooth finish

### Sterilization

Non-Sterile

### Recommended for:

- Controlled environments
- Pharmaceutical manufacturing
- Medical device manufacturing

## Shelf life and storage

- Three years from date of manufacture
- Store in unopened cartons, away from excessive cold/heat, ozone and ultraviolet light
- Preferred temperature range: >40°F (4.44°C) and <100°F (37.77°C)

To obtain certificates of conformance, visit [certificates.cardinalhealth.com](https://certificates.cardinalhealth.com) or email [cleanroomgloves@cardinalhealth.com](mailto:cleanroomgloves@cardinalhealth.com)

Physical properties		
Property	Typical value	Test method
<b>Length</b>	12 inch / 305 mm	ASTM D3577
<b>Thickness</b>	Fingertip 8.3 mil / 0.21 mm	
	Palm 7.1 mil / 0.18 mm	
	Cuff 6.0 mil / 0.15 mm	
<b>Tensile Strength</b>	Before aging Min. 24 MPa	ASTM D412
	After aging Min. 18 MPa	
<b>Ultimate Elongation</b>	Before aging Min. 750%	
	After aging Min. 560%	
<b>Stress at 500% Elongation</b>	Before aging Max. 5.5 MPa	
<b>Freedom from Holes</b>	AQL 0.65	ISO 2859-1

Glove cleanliness		
Extractable Ions (mg/cm <sup>2</sup> )	Typical value	Test method
Bromide	0.00	CCLT-5010/ WI-QA-066
Calcium	1.40	
Chloride	2.99	
Fluoride	0.00	
Magnesium	0.02	
Nitrate	0.44	
Nitrite	0.00	
Phosphate	0.00	
Potassium	0.03	
Sodium	0.28	
Sulfate	0.57	

Quality standards	
<b>Audit Standards</b>	Manufactured in an ISO 13485 certified production facility
<b>Chemical Resistance</b>	Tested against permeation standard ASTM F739 for use with common disinfectants and cleaning agents

**For more information or to request certificates, please contact [cleanroomgloves@cardinalhealth.com](mailto:cleanroomgloves@cardinalhealth.com). To order, contact your local Cardinal Health or distributor sales representative.**

<sup>1</sup>Bennett, M.A. and Tekamp, D.A. (2012). Surgical Glove Comparison Testing.

**Caution: This Product Contains Natural Rubber Latex Which May Cause Allergic Reactions.**

