

Think Pink Once A Week®



COMBINATION PACKAGE

LIQUID ULTRA SOLUTION 1 and LIQUID ULTRA SOLUTION 2

Liquid Ultra Solution 1

This is part one of a two-part product.
Always Use With Liquid Ultra Solution 2

Active Ingredient:

Hydrogen Peroxide	1.05%
Other Ingredients	98.95%
TOTAL.....	100.00%

KEEP OUT OF THE REACH OF CHILDREN DANGER

PRECAUTIONARY STATEMENTS

Hazards To Humans and Domestic Animals. DANGER.

Corrosive. Causes irreversible eye damage. Harmful if absorbed through skin. Do not get in eyes, on skin, or on clothing. Wear protective eyewear (goggles, face shield, or safety glasses) and rubber gloves when handling. Wash thoroughly with soap and water after handling and before eating, drinking, chewing gum, or using tobacco. Remove and wash contaminated clothing before reuse.

FIRST AID

Have the product container or label with you when calling a poison control center or doctor, or going for treatment.

IF IN EYES: Hold eye open and rinse slowly and gently with water for 15-20 minutes. Remove contact lenses, if present, after the first 5 minutes, then continue rinsing. Call a poison control center or doctor for treatment advice.

IF ON SKIN OR CLOTHING: Take off contaminated clothing. Rinse skin immediately with plenty of water for 15-20 minutes. Call a poison control center or doctor for treatment advice.

EPA REG. NO. 74245-7

EPA EST. NO. 73516-MD-001

Rev. 001

Liquid Ultra Solution 2

This is part two of a two-part product.
Always Use With Liquid Ultra Solution 1

Active Ingredients:

n-Alkyl(C ₁₂ 68%, C ₁₄ 32%) dimethyl ethylbenzyl ammonium chloride	0.5%
n-Alkyl(C ₁₄ 60%, C ₁₆ 30%, C ₁₂ 5%, C ₁₈ 5%) dimethyl benzyl ammonium chloride	0.5%
Other Ingredients	99.0%
TOTAL.....	100.0%

KEEP OUT OF THE REACH OF CHILDREN CAUTION

PRECAUTIONARY STATEMENTS

Hazards To Humans and Domestic Animals. CAUTION.

Causes moderate eye irritation and harmful if swallowed. Avoid contact with eyes, skin or clothing. Do not get in eyes, on skin or on clothing. Wash thoroughly with soap and water after handling and before eating, drinking, chewing gum, or using tobacco.

FIRST AID

Have the product container or label with you when calling a poison control center or doctor, or going for treatment.

IF IN EYES: Hold eye open and rinse slowly and gently with water for 15-20 minutes. Remove contact lenses, if present, after the first 5 minutes, then continue rinsing. Call a poison control center or doctor for treatment advice.

IF ON SKIN OR CLOTHING: Take off contaminated clothing. Rinse skin immediately with plenty of water for 15-20 minutes. Call a poison control center or doctor immediately for treatment advice.

IF SWALLOWED: Call a poison control center or doctor immediately for treatment advice. Do not induce vomiting or give anything by mouth to an unconscious person. Drink promptly a large quantity of milk, egg whites, gelatin, or if these are not available, drink large quantities of water. Avoid alcohol.

IF INHALED: Move person to fresh air. If person is not breathing, call 911 or an ambulance, then give artificial respiration, preferably by mouth-to-mouth, if possible. Call a poison control center or doctor for further treatment advice.

EPA REG. NO. 74245-6

EPA EST. NO. 73516-MD-001

Rev. 001

STORAGE AND DISPOSAL

Do not contaminate water, food or feed by improper storage or disposal.

PESTICIDE STORAGE: Keep product in closed original container when not in use. Store in a cool, dry area away from potential sources of heat, open flames, sunlight or other chemicals. DO NOT store above 30°C (or 86°F).

PESTICIDE DISPOSAL: Pesticide wastes are acutely hazardous. Improper disposal of excess pesticide or rinsate is a violation of Federal Law. If these wastes cannot be disposed of by use according to label instructions, contact your State Pesticide or Environmental Control Agency, or the Hazardous Waste Representative at the nearest EPA Regional Office for guidance.

CONTAINER DISPOSAL: Nonrefillable container. Do not reuse or refill this container. Offer for recycling, if available or puncture and dispose of in a sanitary landfill, or by other procedures approved by state and local authorities. If rinsate cannot be used, follow pesticide disposal instructions. If not triple rinsed these containers are acute hazardous wastes and must be disposed in accordance with local, state and federal regulations. Triple rinse container (or equivalent) promptly after emptying. Triple rinse as follows: Empty the remaining contents into application equipment or a mix tank and drain for 10 seconds after the flow begins to drip. Fill the container ¼ full with water and recap. Shake for 10 seconds. Pour rinsate into application equipment or mix tank or store rinsate for later use or disposal. Drain for 10 seconds after the flow begins to drip. Repeat this procedure two more times.

Important Notice:

Before using Crosstex Liquid Ultra Solutions 1 and 2 please read the following information.

1. This product is to be used in conjunction with regular testing of dental unit water. Testing frequency must comply with user's practice protocol.
2. The manufacturer of the dental unit should be consulted before use of this product regarding compatibility of the Liquid Ultra Solution with the dental unit.

DIRECTIONS FOR USE:

It is a violation of Federal law to use this product in a manner inconsistent with its labeling.

Always Use Crosstex Liquid Ultra Solution 1 with Crosstex Liquid Ultra Solution 2.

1. Initial Start-Up Treatment: Treat the Dental Unit Water Lines (DUWL) for three consecutive nights.
2. Routine Treatment: Treat DUWL one night/week (Monday-Thursday recommended) after Initial Start-up treatment.
3. If two or more weekly treatments are skipped, repeat initial start-up treatment.

NOTE: Do Not Use Thin Walled Polyethylene Terephthalate (PET) Bottles. We Recommend Either A High Density Polyethylene Bottle With A Minimum Thickness of 0.08 Inches or A High/Low Density Polyethylene Blend With A Minimum Thickness of 0.14 Inches.

HOW TO APPLY CROSSTEX LIQUID ULTRA SOLUTIONS 1 & 2:

1. Add one bottle (3 ounces) of Crosstex Liquid Ultra Solution 1 and one bottle (3 ounces) of Crosstex Liquid Ultra Solution 2 into an empty external dental unit water bottle and stir. Do not use PET bottles (see note above). Once mixed together, solution must be used within 24 hours.
2. Run Liquid Ultra mixture through the system until the pink solution appears at the end of each air/water syringe and handpiece lines. Depending on the number of lines in the unit, there may be residual solution in the bottle. Always remove the handpiece. It may be advisable to remove coupler depending on the type of coupler, contact manufacturer for specific recommendations.
3. Allow the Liquid Ultra mixture to remain in the lines overnight. To prevent accidental spillage, place the ends of water lines into a sink or container in case any pink mixture drips overnight.
NOTE: Do not allow the Liquid Ultra mixture to remain in the lines for more than 24 hours.
4. The following morning, flush any remaining Liquid Ultra mixture from the bottle, through either the air/water syringe or handpiece lines until the external water bottle is empty.
5. Remove and rinse the external water bottle with water and then fill with water.
6. Flush each line (air/water syringe and handpiece lines) until the water runs clear and the bottle is empty.

For Technical Assistance Call **1-888-276-7783**

CROSSTEX[®]
A CANTEL MEDICAL COMPANY



Manufactured for: Crosstex International, Inc.
Corporate Office: 10 Ranick Road, Hauppauge, NY 11788-4209 USA
Website: www.crosstex.com E-mail: crosstex@crosstex.com

Made in USA

Net Contents: 60 fl. oz.

Crosstex[®] and Liquid Ultra[™] are trademarks and registered trademarks of Crosstex International, Inc.

DS503L-IFU
Rev. 0 - 5/17